

Fast Thermal Conductivity Analyzer FTC300



- Accurate and long term stable thermal conductivity measurement
- High sensitivity e.g. 0-0.5Vol.% H₂ in N₂ range; Low noise <10ppm H₂ in N₂
- Precise linearization for the gas mixtures as H₂, He, CO₂, CH₄ in N₂ or Ar and N₂ in Ar on board
- Customized linearization with polynomial up to sixth order
- Multi Gas Mode allows sequential indication of up to 15 binary mixtures
- Isolated 4-20mA output, expandable; free set of start and end within 100 Vol.% range
- Classic 2-point calibration or simple one-gas calibration
- Free set of display indication (ppm or Vol.%) at a resolution up to 1ppm
- Cross sensitivity compensation by feed in of external signal and internal calculation
- Fast response with a T90-time of less than 1 sec (depending on flow rate)
- Pressure proof (20bar) and vacuum leak tight stainless steel (LF316i) gas duct
- Three isolated relays for indication of alarms and instrument status
- RS 232 excess to all values and parameters
- Digital output with 1ppm resolution over the whole 100Vol.% range

Messkonzept GmbH Analytical Technology Niedwiesenstr. 33 60431 Frankfurt Germany Fon +49 69 53056444 Fax +49 69 53056445

info@messkonzept.de www.messkonzept.de Geschäftsführer Dr. Axel-Ulrich Grunewald Gerichtsstand Frankfurt HRB 49940 USt-ID: DE211207233 Frankfurter Volksbank Konto: 7000903005 BLZ: 50190000 Swift-BIC: FFVBDEFF IBAN: DE03501900007000903005



- PC-based service program simplifies all settings, linearization and calibration
- Small and robust transmitter in Al-housing for field use (protection class: IP65)
- Dimensions: Width 145mm, Height 80mm (without connectors), Depth 85mm
- Power supply 18V to 36V DC / 700mA

Measuring Multi Gas Carrier Gas Basic range Smallest range Smallest suppressed Gas zero range Mode 0% - 100% H_2 N_2 or air 0% - 0.5% 98% - 100% Yes 0% - 100% 0% - 0.4% 99% - 100% H_2 Ar Yes H_2 He 20% - 100% 20% - 40% 85% - 100% On request H_2 CH₄ 0% - 100% 0% - 0.5% 98% - 100% On request H_2 CO_2 0% - 100% 0% - 0.5% 98% - 100% On request 0% - 100% 97% - 100% He N_2 or air 0% - 0.8% Yes He Ar 0% - 100% 0% - 0.5% 98% - 100% Yes 0% - 100% CO₂ N_2 or air 0% - 3% 96% - 100% Yes CO2 0% - 60% 0% - 10% Ar Yes 0% - 3% Ar N_2 or air 0% - 100% 96% - 100% Yes 0% - 100% 0% - 3% Ar 02 96% - 100% Yes 40% - 100% 80% - 100% Ar CO_2 -Yes CH₄ N_2 or air 0% - 100% 0% - 2% 96% - 100% Yes 0% - 100% 0% - 1.5% 97% - 100% Yes CH₄ Ar 0% - 100% 0% - 15% 85% - 100% Yes 02 N_2 Ar 0% - 100% 0% - 2% 97% - 100% Yes 02 N_2 0% - 100% 0% - 2% 99,5% - 100% Yes H_2 N_2 Ar 0% - 100% 0% - 3% 97% - 100% Yes 96% - 100% 0% - 100% 0% - 4% N_2 CO, On request NH₃ H_2 0% - 100% 0% - 5% 95% - 100% On request NH₃ 0% - 100% 0% - 10% 60% - 100% On request N_2 CO H_2 0% - 100% 0% - 2% 99% - 100% On request 0% - 100% 0% - 2% 96% - 100% SF_6 N_2 or air On request

Measuring Ranges:

Other gases and ranges on request; Multi Gas Mode "Yes" means that these binary mixtures and in addition one customized gas mixture may be measured sequentially with one instrument

Specification:

Dimensions without connectors; weight	145mm x 80mm x 85mm; max. 1800g
Power supply	24V DC (18V to 36V), 700mA
Ambient temperature range	-5°C to 50°C, other on request
Linearity	< 1% of range
Warm up time	About 30min; 1h for small ranges
Flow rate	40I/h to 150I/h; 60I/h -80I/h recommended
T90-time	<1sec at flow rate higher 60l/h
Noise	< 1% of smallest range
Drift at zero point	< 2% of smallest range per week
Repeatability	< 1% of range
Error due to change of ambient temperature	< 1% of smallest range per 10°C
Error due to change of flow at 801/h	< 1% of smallest range per 10l/h
Gas pressure	Max. 2MPa (20bar)
Error due to change of pressure (above 800hPa abs.)	< 1% of smallest range per 10hPa

Important Notice: The specifications are given for guidance; they might differ for some gas mixture.

Messkonzept GmbH Analytical Technology Niedwiesenstr. 33 60431 Frankfurt Germany Fon + 49 69 53056444 Fax + 49 69 53056445 info@messkonzept.de Geschäftsführer Dr. Axel-Ulrich Grunewald Gerichtsstand Frankfurt HRB 49940 USt-ID: DE211207233 Frankfurter Volksbank Konto: 7000903005 BLZ: 50190000 Swift-BIC: FFVBDEFF IBAN: DE03501900007000903005